

UNITED STATES DEPARTMENT OF COMMERCE United States Patent and Trademark Office Addiese: COMMISSIONER FOR PATENTS P O Box 1450 Alexandra, Virginia 22313-1450 www.wepto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/531,860	11/16/2005	Francois Giordano	12400-037	3997
757 BRINKS HOE	757 7590 04/06/2009 BRINKS HOFER GILSON & LIONE		EXAMINER	
P.O. BOX 103	95		PATTON, SPENCER D	
CHICAGO, IL 60610			ART UNIT	PAPER NUMBER
			3664	•
			MAIL DATE	DELIVERY MODE
			04/06/2009	PAPER

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

## Advisory Action Before the Filing of an Appeal Brief

	Application No.	Applicant(s)				
	10/531,860	GIORDANO, FRANCOIS				
	Examiner	Art Unit				
	SPENCER PATTON	3664				

--The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

THE REPLY FILED 23 March 2009 FAILS TO PLACE THIS APPLICATION IN CONDITION FOR ALLOWANCE.

- 1. X The reply was filed after a final rejection, but prior to or on the same day as filing a Notice of Appeal. To avoid abandonment of this application, applicant must timely file one of the following replies: (1) an amendment, affidavit, or other evidence, which places the application in condition for allowance; (2) a Notice of Appeal (with appeal fee) in compliance with 37 CFR 41.31; or (3) a Request for Continued Examination (RCE) in compliance with 37 CFR 1.114. The reply must be filed within one of the following time periods:
  - a) The period for reply expires 3 months from the mailing date of the final rejection.
  - The period for reply expires on: (1) the mailing date of this Advisory Action, or (2) the date set forth in the final rejection, whichever is later. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the mailing date of the final rejection.

Examiner Note: If box 1 is checked, check either box (a) or (b). ONLY CHECK BOX (b) WHEN THE FIRST REPLY WAS FILED WITHIN TWO MONTHS OF THE FINAL REJECTION. See MPEP 706.07(f).

Extensions of time may be obtained under 37 CFR 1.136(a). The date on which the petition under 37 CFR 1.136(a) and the appropriate extension fee have been filed is the date for purposes of determining the period of extension and the corresponding amount of the fee. The appropriate extension fee under 37 CFR 1.17(a) is calculated from: (1) the expiration date of the shortened statutory period for reply originally set in the final Office action; or (2) as set forth in (b) above, if checked. Any reply received by the Office later than three months after the mailing date of the final rejection, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

## NOTICE OF APPEAL

2. The Notice of Appeal was filed on ... A brief in compliance with 37 CFR 41.37 must be filed within two months of the date of filing the Notice of Appeal (37 CFR 41.37(a)), or any extension thereof (37 CFR 41.37(e)), to avoid dismissal of the appeal. Since a Notice of Appeal has been filed, any reply must be filed within the time period set forth in 37 CFR 41.37(a).

## AMENDMENTS

- 3. 🔯 The proposed amendment(s) filed after a final rejection, but prior to the date of filing a brief, will not be entered because

  - (a) ☐ They raise new issues that would require further consideration and/or search (see NOTE below);
    (b) ☐ They raise the issue of new matter (see NOTE below); (c) They are not deemed to place the application in better form for appeal by materially reducing or simplifying the issues for
    - appeal; and/or
    - (d) They present additional claims without canceling a corresponding number of finally rejected claims.
- NOTE: See Continuation Sheet. (See 37 CFR 1.116 and 41.33(a)).
- The amendments are not in compliance with 37 CFR 1.121. See attached Notice of Non-Compliant Amendment (PTOL-324).
- Applicant's reply has overcome the following rejection(s): 6. Newly proposed or amended claim(s) would be allowable if submitted in a separate, timely filed amendment canceling the non-allowable claim(s).
- 7. To purposes of appeal, the proposed amendment(s): a) will not be entered, or b) will be entered and an explanation of how the new or amended claims would be rejected is provided below or appended.
  - The status of the claim(s) is (or will be) as follows:
  - Claim(s) allowed:
  - Claim(s) objected to: \_\_\_ Claim(s) rejected: \_
  - Claim(s) withdrawn from consideration: \_\_\_

## AFFIDAVIT OR OTHER EVIDENCE

- 8. The affidavit or other evidence filed after a final action, but before or on the date of filing a Notice of Appeal will not be entered because applicant failed to provide a showing of good and sufficient reasons why the affidavit or other evidence is necessary and was not earlier presented. See 37 CFR 1.116(e).
- 9. The affidavit or other evidence filed after the date of filing a Notice of Appeal, but prior to the date of filing a brief, will not be entered because the affidavit or other evidence failed to overcome all rejections under appeal and/or appellant fails to provide a showing a good and sufficient reasons why it is necessary and was not earlier presented. See 37 CFR 41.33(d)(1).
- 10. The affidavit or other evidence is entered. An explanation of the status of the claims after entry is below or attached.
- REQUEST FOR RECONSIDERATION/OTHER
- 11. The request for reconsideration has been considered but does NOT place the application in condition for allowance because:
- 12. Note the attached Information Disclosure Statement(s). (PTO/SB/08) Paper No(s).
- 13. Other: The arguments related to the additional limitation added to claim 1: "the sensor unit has a smaller volume than that of the control unit" are not persuasive for the reasons mentioned in the Office action mailed 1/21/2009 with respect to claim 17. These reasons can be summarized as: McCall (6,522,992) teaches a small sensor unit, and MPEP 2144.04 (IV) (A) renders the specific relation of volume of the control unit to volume of the sensor unit obvious since the control unit being larger than the sensor unit is not critical and does not produce any unexpected results. The advantages listed by applicant on page 8 of the arguments mailed 3/23/2009 all result from the sensor unit being separate from the control unit, and from the sensor unit being small, both of which are taught by the prior art of record.

Continuation Sheet (PTOL-303)

Application No.

/KHOI TRAN/ Supervisory Patent Examiner, Art Unit 3664 /Spencer Patton/

U.S. Patent and Trademark Office

PTOL-303 (Rev. 08-06)

Advisory Action Before the Filing of an Appeal Brief

Part of Paper No. 20090330

Continuation of 3. NOTE: The additional limitation of "a control unit located... away from the central longitudinal axis" in claim 24 raises new issues that would require further consideration. Additionally the proposed amendment to claim 1 does not avoid any of the rejections set forth in the previous Office action, and thus the amendment would not place the case in condition for allowance.